

**Do I need to re-measure my standards that were measured with a Datacolor 600 or 800?**

No, standards do not need to be re-measured. Existing standards measured on other Datacolor spectrophotometers can be imported and used along with SpectraVision  $\hat{A}$   $\hat{A}$   $\hat{A}$

*Author:* Datacolor

*Saved From:* <http://knowledgebase.datacolor.com/article-686.html>