Do I need to re-measure my standards that were measured with a Datacolor 600 or 800?

No, standards do not need to be re-measured. Existing standards measured on other Datacolor spectrophotometers can be imported and used along with SpectraVision Â Â Â

Author: Datacolor
Saved From: http://knowledgebase.datacolor.com/article-686.html