

**I still receive a calibration error after correct calibration file data is installed.**

If the instrument is communicating properly and you are getting a message "*instrument not calibrated*" after completing calibration, open c:\Program files\Datacolor\Common Files\CometFlashClientTest.exe and set calibration interval to 8 hours by selecting Calibrate Service button, and Calibrate tab. Change Calibration Time Interval (hours) to 8.

Author: Datacolor

Saved From: <http://knowledgebase.datacolor.com/article-137.html>