

Why are there always 2 possibilities for one illuminant?

Each illuminant is defined for 2 standardized observers, which can be 2° or 10°. The angles are related to the different sizes of the observer's field of view.

Author: Datacolor

Saved From: <http://knowledgebase.datacolor.com/article-509.html>