

**Green tile test gives unstable results when TOOLS is connected to Microflash 45 IR (MF45 IR)**

Answer: **Turn off the AVG mode** of MF45 IR.

When the AVG is on, the instrument calculates the average of the last measures and sends it to TOOLS.

*Author: Datacolor*

*Saved From: <http://knowledgebase.datacolor.com/article-618.html>*